
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001720
15:06 21/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:21
136.0 C 221kPa 3:21
136.3 C 224kPa 3:51
136.4 C 225kPa 4:21
136.3 C 225kPa 4:51
136.4 C 224kPa 5:21
136.3 C 225kPa 5:51
136.4 C 225kPa 6:21
136.5 C 223kPa 6:51
136.5 C 223kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 37:43

Digital Signature #
4D47F2885337676D
