

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000293  
11:41 06/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:46  
STERILIZING 4:14  
136.4 C 230kPa 4:14  
136.3 C 232kPa 4:44  
136.3 C 232kPa 5:14  
136.4 C 232kPa 5:44  
136.3 C 229kPa 6:14  
136.4 C 232kPa 6:44  
136.3 C 231kPa 7:14  
136.4 C 232kPa 7:44  
136.4 C 232kPa 7:44  
VENTING 7:44  
AIR DRYING 8:38  
CYCLE COMPLETE 10:53  
-----

Digital Signature #  
09C162B18851F3D3  
-----