
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000058
10:48 26/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:19
STERILIZING 4:47
136.4 C 224kPa 4:47
135.8 C 221kPa 5:17
136.4 C 225kPa 5:47
136.4 C 225kPa 6:17
136.4 C 224kPa 6:47
136.4 C 224kPa 7:17
136.4 C 221kPa 7:47
136.4 C 223kPa 8:17
136.4 C 223kPa 8:17
VENTING 8:17
AIR DRYING 9:07
CYCLE COMPLETE 18:43

Digital Signature #
B65201CD2B721AED
