

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1827

Name: RITTER2  
Sterilizer ID: V2313228  
12/08/2022 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.9 F 0.0  
2:00 148.5 F 2.6  
4:00 168.5 F 6.3  
6:00 196.2 F 11.1  
8:00 214.9 F 4.9  
10:00 226.9 F 10.1  
12:00 231.3 F 7.9  
14:00 242.7 F 13.0  
16:00 242.1 F 11.8  
18:00 255.1 F 18.5  
20:00 266.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.6  
2:00 271.5 F 28.9  
3:00 272.5 F 29.8  
4:00 272.1 F 29.3

Min 271.4 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 214.8 F 0.0  
10:00 228.1 F 0.0  
15:00 229.3 F 0.0  
20:00 227.6 F 0.0  
25:00 225.8 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 21:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

12/08/2022 03:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_