

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5178

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2024 10:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.4 F 0.0  
2:00 168.8 F 0.0  
4:00 181.1 F 0.2  
6:00 200.7 F 3.6  
8:00 217.9 F 8.1  
10:00 229.8 F 12.9  
12:00 232.2 F 8.7  
14:00 240.5 F 11.2  
16:00 243.1 F 12.1  
18:00 240.8 F 10.7  
20:00 254.3 F 17.5  
22:00 265.7 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 28.9  
2:00 271.5 F 28.1  
3:00 272.7 F 29.0  
4:00 272.1 F 28.5

Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.8 F 0.0  
5:00 210.2 F 0.0  
10:00 213.7 F 0.0  
15:00 221.0 F 0.0  
20:00 227.0 F 0.0  
25:00 229.8 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 23:44  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:01:50  
CYCLE COMPLETE

12/11/2024 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_