
Midmark M11 - v1.0.0
Total Cycles: 4651

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2024 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 76.0 F 0.0
4:00 86.3 F 0.1
6:00 119.5 F 0.2
8:00 154.8 F 1.1
10:00 171.6 F 4.4
12:00 190.3 F 7.8
14:00 205.8 F 11.3
16:00 214.9 F 4.6
18:00 225.2 F 8.1
20:00 235.1 F 12.2
22:00 233.6 F 8.3
24:00 244.5 F 12.7
26:00 243.5 F 11.9
28:00 255.3 F 17.9
30:00 265.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.5 F 28.3
3:00 272.3 F 28.6
4:00 271.4 F 28.3

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 220.9 F 0.0
10:00 216.3 F 0.0
15:00 221.9 F 0.0
20:00 233.2 F 0.0
25:00 226.0 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:01
CYCLE COMPLETE

06/18/2024 09:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____