
Midmark M11 - v1.0.5
Total Cycles: 4374

Name: RITTER2
Sterilizer ID: V2313228
02/24/2025 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.3 F 0.0
2:00 146.0 F 1.9
4:00 158.8 F 5.0
6:00 187.3 F 9.2
8:00 204.2 F 13.5
10:00 212.7 F 6.6
12:00 225.1 F 11.0
14:00 230.0 F 7.2
16:00 241.6 F 12.4
18:00 239.5 F 10.8
20:00 251.3 F 16.7
22:00 262.0 F 22.8
24:00 271.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.7
1:00 272.0 F 29.6
2:00 272.4 F 30.0
3:00 272.2 F 30.2
4:00 271.3 F 29.8

Min 271.2 F 29.6
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 228.4 F 0.0
10:00 218.4 F 0.0
15:00 185.4 F 0.0
20:00 172.3 F 0.0
25:00 168.9 F 0.0
30:00 174.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:04:05
CYCLE COMPLETE

02/24/2025 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____