
Midmark M11 - v1.0.5
Total Cycles: 2474

Name: RITTER2
Sterilizer ID: V2313228
06/06/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.0 F 0.0
2:00 164.6 F 3.2
4:00 183.4 F 7.7
6:00 206.9 F 13.2
8:00 223.4 F 8.4
10:00 233.3 F 10.9
12:00 238.6 F 11.6
14:00 238.2 F 10.3
16:00 251.8 F 17.1
18:00 263.8 F 24.0
20:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.8
2:00 271.2 F 29.1
3:00 272.3 F 30.0
4:00 272.1 F 29.6

Min 271.1 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 217.9 F 0.0
10:00 224.5 F 0.0
15:00 219.2 F 0.0
20:00 222.2 F 0.0
25:00 226.0 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

06/06/2023 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____