
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000371
9:54 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:17
STERILIZING 3:26
136.0 C 226kPa 3:26
136.4 C 233kPa 3:56
136.1 C 232kPa 4:26
136.1 C 230kPa 4:56
136.2 C 230kPa 5:26
136.4 C 233kPa 5:56
136.4 C 233kPa 6:26
136.5 C 233kPa 6:56
136.5 C 233kPa 6:56
VENTING 6:56
AIR DRYING 7:47
CYCLE COMPLETE 7:49

Digital Signature #
ECFE2DEB403286BA
