

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1378

Name: RITTER 1  
Sterilizer ID: V2178198  
11/17/2021 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
2:00 197.0 F 3.7  
4:00 198.1 F 8.8  
6:00 232.4 F 9.0  
8:00 229.9 F 10.4  
10:00 236.3 F 7.7  
12:00 242.7 F 11.0  
14:00 245.6 F 13.4  
16:00 258.4 F 20.2  
18:00 269.8 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.6 F 29.3  
2:00 271.3 F 28.6  
3:00 272.2 F 29.4  
4:00 272.3 F 29.1  
  
Min 271.0 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 231.9 F 0.0  
10:00 227.8 F 0.0  
15:00 227.3 F 0.0  
20:00 226.7 F 0.0  
25:00 228.1 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

11/17/2021 02:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_