

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3846

Name: RITTER2  
Sterilizer ID: V2313228  
08/12/2024 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.5 F 0.0  
2:00 120.7 F 0.6  
4:00 122.3 F 2.4  
6:00 141.5 F 5.1  
8:00 171.5 F 8.4  
10:00 183.9 F 12.1  
12:00 211.5 F 4.7  
14:00 221.2 F 8.4  
16:00 229.7 F 12.4  
18:00 232.6 F 8.8  
20:00 241.8 F 12.4  
22:00 242.3 F 12.1  
24:00 253.4 F 17.7  
26:00 263.2 F 23.6  
28:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.9  
1:00 272.1 F 29.8  
2:00 272.3 F 29.8  
3:00 272.5 F 30.0  
4:00 272.5 F 30.2

Min 271.9 F 29.8  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 216.6 F 0.0  
10:00 208.9 F 0.0  
15:00 200.0 F 0.0  
20:00 202.0 F 0.0  
25:00 206.4 F 0.0  
30:00 202.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 29:10  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:07:48  
CYCLE COMPLETE

08/12/2024 11:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_