
Midmark M11 - v1.0.0
Total Cycles: 2106

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2022 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.3 F 0.0
2:00 132.7 F 0.2
4:00 167.8 F 2.2
6:00 195.3 F 5.6
8:00 216.0 F 10.1
10:00 221.1 F 6.3
12:00 230.1 F 8.4
14:00 240.7 F 12.8
16:00 237.3 F 9.3
18:00 238.8 F 9.5
20:00 246.3 F 13.2
22:00 258.1 F 19.5
24:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 29.1
2:00 271.7 F 28.6
3:00 272.2 F 28.5
4:00 272.6 F 29.0

Min 270.7 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 215.4 F 0.0
10:00 199.0 F 0.0
15:00 204.7 F 0.0
20:00 207.8 F 0.0
25:00 200.1 F 0.0
30:00 202.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

07/08/2022 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____