
Midmark M11 - v1.0.5
Total Cycles: 2308

Name: RITTER2
Sterilizer ID: V2313228
04/27/2023 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.4 F 0.0
2:00 148.3 F 2.3
4:00 162.0 F 5.8
6:00 189.5 F 10.9
8:00 214.3 F 4.9
10:00 225.8 F 10.1
12:00 231.2 F 8.0
14:00 242.6 F 13.2
16:00 242.8 F 12.1
18:00 255.7 F 18.9
20:00 267.1 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 272.2 F 29.7
2:00 271.8 F 29.0
3:00 271.2 F 29.2
4:00 272.2 F 29.7

Min 271.0 F 28.7
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 207.6 F 0.0
10:00 225.5 F 0.0
15:00 229.1 F 0.0
20:00 226.7 F 0.0
25:00 224.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

04/27/2023 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____