

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4782

Name: RITTER 1  
Sterilizer ID: V2178198  
08/01/2024 12:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.6 F 0.0  
2:00 88.4 F 0.1  
4:00 115.9 F 0.7  
6:00 148.9 F 3.0  
8:00 172.5 F 5.9  
10:00 193.7 F 8.9  
12:00 204.5 F 12.2  
14:00 212.2 F 4.4  
16:00 222.2 F 8.1  
18:00 232.6 F 12.2  
20:00 232.4 F 8.2  
22:00 244.1 F 13.4  
24:00 242.2 F 11.5  
26:00 253.6 F 17.3  
28:00 263.7 F 23.2  
30:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.8 F 29.2  
2:00 272.1 F 28.6  
3:00 271.9 F 28.7  
4:00 272.8 F 29.2

Min 271.6 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 214.5 F 0.0  
10:00 221.9 F 0.0  
15:00 220.6 F 0.0  
20:00 220.9 F 0.0  
25:00 222.7 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 30:18  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

08/01/2024 01:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_