

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2468

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2023 07:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.5 F 0.0  
2:00 75.8 F 0.0  
4:00 82.0 F 0.0  
6:00 102.4 F 0.0  
8:00 155.5 F 0.8  
10:00 184.9 F 4.2  
12:00 207.4 F 7.8  
14:00 217.5 F 11.5  
16:00 223.9 F 5.5  
18:00 231.7 F 9.9  
20:00 236.5 F 9.9  
22:00 240.7 F 11.8  
24:00 238.6 F 10.1  
26:00 250.9 F 16.2  
28:00 261.6 F 22.4  
30:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.5 F 29.5  
2:00 271.5 F 29.2  
3:00 272.3 F 29.4  
4:00 271.4 F 29.2

Min 271.2 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 228.3 F 0.0  
10:00 217.4 F 0.0  
15:00 216.5 F 0.0  
20:00 206.9 F 0.0  
25:00 223.5 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 30:36  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:33  
CYCLE COMPLETE

06/06/2023 08:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_