

-----  
Midmark M11 - v1.0.0  
Total Cycles: 922

Name: RITTER 1  
Sterilizer ID: V2178198  
06/22/2021 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 74.2 F 0.0  
4:00 104.6 F 0.0  
6:00 122.2 F 0.1  
8:00 145.6 F 0.1  
10:00 173.2 F 0.2  
12:00 193.7 F 0.6  
14:00 208.0 F 3.3  
16:00 214.8 F 5.6  
18:00 222.7 F 8.3  
20:00 231.5 F 11.6  
22:00 229.1 F 6.6  
24:00 240.2 F 11.6  
26:00 237.0 F 9.1  
28:00 240.9 F 10.3  
30:00 247.4 F 13.8  
32:00 258.8 F 20.0  
34:00 269.1 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.7 F 28.8  
2:00 271.6 F 28.0  
3:00 272.5 F 28.9  
4:00 272.1 F 28.4  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 188.6 F 0.0  
10:00 186.7 F 0.0  
15:00 190.0 F 0.0  
20:00 190.0 F 0.0  
25:00 192.3 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 35:09  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:13:28  
CYCLE COMPLETE

06/22/2021 08:00 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_