
Midmark M11 - v1.0.0
Total Cycles: 2057

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 07:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 75.3 F 0.0
4:00 97.3 F 0.0
6:00 109.6 F 0.8
8:00 145.6 F 2.4
10:00 173.4 F 4.7
12:00 191.2 F 7.6
14:00 203.8 F 11.0
16:00 213.2 F 4.9
18:00 222.5 F 7.2
20:00 232.8 F 11.1
22:00 231.2 F 7.0
24:00 243.2 F 12.6
26:00 240.5 F 10.6
28:00 253.0 F 16.9
30:00 263.5 F 23.1
32:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.2
2:00 271.8 F 28.4
3:00 271.9 F 28.8
4:00 272.6 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.6 F 0.0
5:00 192.1 F 0.0
10:00 205.0 F 0.0
15:00 196.6 F 0.0
20:00 215.3 F 0.0
25:00 230.8 F 0.0
30:00 235.7 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:12
CYCLE COMPLETE

06/27/2022 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____