
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005030
14:51 31/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:04
STERILIZING 4:14
136.5 C 221kPa 4:14
136.4 C 225kPa 4:44
136.4 C 223kPa 5:14
136.4 C 224kPa 5:44
136.4 C 220kPa 6:14
136.3 C 224kPa 6:44
136.4 C 224kPa 7:14
136.4 C 224kPa 7:44
136.4 C 224kPa 7:44
VENTING 7:44
AIR DRYING 8:44
CYCLE COMPLETE 8:50

Digital Signature #
B2F3E6AEC676F264
