
Midmark M11 - v1.0.0
Total Cycles: 1793

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2022 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 163.2 F 2.4
4:00 173.1 F 5.6
6:00 193.6 F 10.2
8:00 216.2 F 4.6
10:00 224.2 F 8.6
12:00 237.5 F 12.4
14:00 238.8 F 10.5
16:00 235.1 F 8.2
18:00 249.9 F 15.2
20:00 261.9 F 22.1
22:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.0
2:00 272.3 F 29.2
3:00 272.1 F 28.6
4:00 272.5 F 29.4

Min 271.4 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 239.7 F 0.0
10:00 228.3 F 0.0
15:00 227.8 F 0.0
20:00 233.0 F 0.0
25:00 223.6 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:60
CYCLE COMPLETE

04/05/2022 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____