
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002480
12:28 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:59
STERILIZING 3:05
136.0 C 222kPa 3:05
135.9 C 222kPa 3:35
136.4 C 225kPa 4:05
136.4 C 225kPa 4:35
136.2 C 224kPa 5:05
136.1 C 223kPa 5:35
136.4 C 226kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 6:35
VENTING 6:35
AIR DRYING 7:29
CYCLE COMPLETE 37:29

Digital Signature #
35493CC10BB0F002
