
Midmark M11 - v1.0.0
Total Cycles: 4383

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2024 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.9 F 0.0
2:00 105.4 F 0.7
4:00 160.3 F 2.7
6:00 157.8 F 5.5
8:00 187.2 F 8.9
10:00 220.7 F 12.4
12:00 217.2 F 3.9
14:00 225.5 F 7.3
16:00 234.3 F 11.0
18:00 231.8 F 7.0
20:00 240.7 F 11.2
22:00 235.7 F 8.5
24:00 248.3 F 14.3
26:00 258.7 F 19.9
28:00 268.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 272.2 F 29.1
3:00 271.5 F 28.6
4:00 272.1 F 28.6

Min 270.9 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.8 F 0.0
5:00 234.0 F 0.0
10:00 201.6 F 0.0
15:00 212.7 F 0.0
20:00 214.0 F 0.0
25:00 222.5 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:13
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:07:58
CYCLE COMPLETE

03/15/2024 03:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____