
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001261
9:46 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
PRESSURIZING 4:20
STERILIZING 5:36
136.0 C 221kPa 5:36
136.0 C 221kPa 6:06
136.2 C 223kPa 6:36
136.3 C 223kPa 7:06
136.4 C 224kPa 7:36
136.4 C 222kPa 8:06
136.2 C 223kPa 8:36
136.2 C 223kPa 9:06
136.2 C 223kPa 9:06
VENTING 9:06
AIR DRYING 10:01
CYCLE COMPLETE 21:36

Digital Signature #
8A0A4D9A9842D099
