
Midmark M11 - v1.0.0
Total Cycles: 1361

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2021 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.6 F 0.0
2:00 193.2 F 3.8
4:00 197.4 F 8.8
6:00 215.5 F 8.3
8:00 232.1 F 8.7
10:00 240.6 F 10.9
12:00 242.8 F 12.0
14:00 241.4 F 10.8
16:00 255.2 F 17.8
18:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.5 F 28.5
2:00 272.1 F 28.4
3:00 272.4 F 28.4
4:00 271.8 F 28.2

Min 271.6 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 217.7 F 0.0
10:00 202.9 F 0.0
15:00 209.5 F 0.0
20:00 219.9 F 0.0
25:00 227.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

11/11/2021 12:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____