
Midmark M11 - v1.0.5
Total Cycles: 2363

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 78.6 F 0.0
4:00 111.5 F 0.0
6:00 108.0 F 0.3
8:00 146.2 F 0.3
10:00 171.4 F 1.1
12:00 196.4 F 4.5
14:00 209.9 F 8.2
16:00 220.8 F 12.1
18:00 225.6 F 7.2
20:00 237.2 F 12.0
22:00 235.4 F 9.3
24:00 238.5 F 9.9
26:00 246.8 F 13.9
28:00 257.9 F 20.0
30:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.6
2:00 271.6 F 28.7
3:00 272.5 F 29.7
4:00 272.2 F 29.1

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 196.3 F 0.0
10:00 205.6 F 0.0
15:00 200.9 F 0.0
20:00 197.4 F 0.0
25:00 199.9 F 0.0
30:00 206.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

05/15/2023 08:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____