
Midmark M11 - v1.0.0
Total Cycles: 3109

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 187.2 F 3.9
4:00 201.5 F 8.6
6:00 222.8 F 8.4
8:00 231.2 F 9.1
10:00 238.6 F 10.2
12:00 241.5 F 12.0
14:00 239.8 F 10.5
16:00 253.2 F 17.1
18:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.6 F 29.2
2:00 271.5 F 28.4
3:00 272.6 F 29.2
4:00 271.6 F 28.5

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 225.7 F 0.0
10:00 225.7 F 0.0
15:00 221.9 F 0.0
20:00 222.4 F 0.0
25:00 222.9 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

04/13/2023 03:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____