
Midmark M11 - v1.0.5
Total Cycles: 929

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.3 F 0.0
2:00 171.1 F 0.1
4:00 197.3 F 2.3
6:00 216.3 F 7.1
8:00 232.8 F 12.2
10:00 233.7 F 8.8
12:00 238.8 F 10.7
14:00 244.7 F 13.0
16:00 244.1 F 12.3
18:00 256.9 F 19.2
20:00 268.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.4 F 29.1
2:00 272.6 F 29.5
3:00 272.2 F 29.7
4:00 272.7 F 29.3

Min 271.0 F 28.4
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 222.9 F 0.0
10:00 234.6 F 0.0
15:00 231.5 F 0.0
20:00 228.5 F 0.0
25:00 227.6 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

03/30/2022 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____