
Midmark M11 - v1.0.5
Total Cycles: 643

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.5 F 0.0
2:00 181.5 F 3.6
4:00 196.2 F 8.9
6:00 213.3 F 7.5
8:00 228.4 F 9.9
10:00 234.0 F 8.4
12:00 242.6 F 12.3
14:00 244.7 F 13.3
16:00 257.8 F 20.4
18:00 269.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.6 F 29.8
2:00 271.8 F 29.5
3:00 272.7 F 29.9
4:00 271.7 F 29.2

Min 271.7 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 205.6 F 0.0
10:00 216.0 F 0.0
15:00 219.5 F 0.0
20:00 217.8 F 0.0
25:00 215.7 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

12/16/2021 12:22 PM
ACCEPTED: X
REJECTED:
Operator: U02