

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006080  
8:31 09/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 3:39  
STERILIZING 4:41  
136.4 C 221kPa 4:41  
136.3 C 222kPa 5:11  
135.8 C 219kPa 5:41  
136.3 C 223kPa 6:11  
136.4 C 223kPa 6:41  
136.4 C 222kPa 7:11  
136.1 C 221kPa 7:41  
136.3 C 223kPa 8:11  
136.3 C 223kPa 8:11  
VENTING 8:11  
AIR DRYING 8:57  
CYCLE COMPLETE 9:26

-----  
Digital Signature #  
41F7391510047679  
-----