
Midmark M11 - v1.0.0
Total Cycles: 5190

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2024 05:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 185.4 F 0.4
4:00 208.8 F 5.7
6:00 225.5 F 11.8
8:00 230.9 F 7.7
10:00 242.3 F 11.9
12:00 244.5 F 12.3
14:00 243.2 F 11.4
16:00 257.8 F 19.0
18:00 270.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.7
1:00 272.1 F 27.9
2:00 272.8 F 28.5
3:00 271.6 F 27.7
4:00 272.5 F 28.2

Min 271.5 F 27.6
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 208.0 F 0.0
10:00 209.3 F 0.0
15:00 214.5 F 0.0
20:00 214.2 F 0.0
25:00 221.5 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

12/13/2024 06:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____