
Midmark M11 - v1.0.5
Total Cycles: 846

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 03:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.5 F 0.0
2:00 164.2 F 3.4
4:00 177.6 F 7.7
6:00 200.4 F 13.1
8:00 222.0 F 7.5
10:00 232.6 F 13.2
12:00 238.0 F 10.9
14:00 237.4 F 9.7
16:00 252.0 F 16.8
18:00 264.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.9 F 30.0
2:00 271.8 F 29.2
3:00 272.8 F 29.9
4:00 271.9 F 29.1

Min 271.0 F 28.7
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 210.8 F 0.0
10:00 218.6 F 0.0
15:00 234.2 F 0.0
20:00 230.6 F 0.0
25:00 228.2 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

02/28/2022 04:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____