

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3918

Name: RITTER 1  
Sterilizer ID: V2178198  
10/18/2023 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.7 F 0.0  
2:00 145.8 F 2.2  
4:00 157.2 F 5.1  
6:00 180.4 F 8.8  
8:00 200.4 F 13.6  
10:00 213.7 F 6.8  
12:00 225.7 F 11.4  
14:00 230.3 F 7.6  
16:00 242.6 F 13.1  
18:00 241.6 F 11.5  
20:00 253.9 F 17.6  
22:00 264.6 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.1 F 29.3  
2:00 272.1 F 28.8  
3:00 271.5 F 28.9  
4:00 272.6 F 29.3

Min 271.2 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.0 F 0.0  
5:00 204.1 F 0.0  
10:00 207.6 F 0.0  
15:00 207.1 F 0.0  
20:00 214.3 F 0.0  
25:00 218.5 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:52  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:02:36  
CYCLE COMPLETE

10/18/2023 11:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_