
Midmark M11 - v1.0.5
Total Cycles: 309

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.1 F 0.0
2:00 193.2 F 4.0
4:00 215.3 F 10.3
6:00 227.0 F 7.2
8:00 239.9 F 11.9
10:00 244.0 F 13.2
12:00 245.0 F 13.1
14:00 258.5 F 20.3
16:00 270.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.6 F 29.2
2:00 272.3 F 29.5
3:00 271.9 F 28.9
4:00 272.4 F 29.6

Min 271.5 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 220.6 F 0.0
10:00 220.5 F 0.0
15:00 214.9 F 0.0
20:00 205.2 F 0.0
25:00 206.1 F 0.0
Dry Cycle Aborted

FILLING: 3:03
HEATING: 16:50
STERILIZING: 4:00
VENTING: 3:03
DRYING: 26:25
TOTAL CYCLE: 00:53:21

08/25/2021 09:41 AM
ACCEPTED: X
REJECTED:
Operator: U02