
Midmark M11 - v1.0.5
Total Cycles: 1967

Name: RITTER2
Sterilizer ID: V2313228
01/24/2023 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 195.0 F 3.9
4:00 207.2 F 9.5
6:00 223.4 F 7.4
8:00 236.9 F 11.6
10:00 237.0 F 10.0
12:00 236.7 F 9.0
14:00 251.6 F 16.6
16:00 264.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.3
2:00 271.8 F 29.2
3:00 272.4 F 29.3
4:00 271.5 F 28.8

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 223.4 F 0.0
10:00 239.0 F 0.0
15:00 230.2 F 0.0
20:00 229.8 F 0.0
25:00 227.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 17:53
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

01/24/2023 03:51 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____