
Midmark M11 - v1.0.0
Total Cycles: 2935

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.8 F 0.0
2:00 81.8 F 0.0
4:00 109.1 F 0.0
6:00 115.7 F 1.8
8:00 140.8 F 4.3
10:00 167.5 F 7.3
12:00 190.4 F 10.6
14:00 212.2 F 4.2
16:00 214.4 F 6.6
18:00 225.4 F 10.3
20:00 233.5 F 9.0
22:00 237.5 F 10.5
24:00 235.3 F 8.0
26:00 246.9 F 13.7
28:00 257.8 F 19.7
30:00 267.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.3 F 29.0
2:00 271.3 F 28.6
3:00 272.2 F 28.9
4:00 271.3 F 28.6

Min 271.2 F 28.3
Max 272.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 205.0 F 0.0
10:00 205.5 F 0.0
15:00 210.7 F 0.0
20:00 215.2 F 0.0
25:00 216.8 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

02/23/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____