
Midmark M11 - v1.0.5
Total Cycles: 4441

Name: RITTER2
Sterilizer ID: V2313228
03/21/2025 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 147.3 F 1.6
4:00 151.4 F 4.4
6:00 170.2 F 8.2
8:00 189.8 F 12.6
10:00 220.5 F 5.5
12:00 227.5 F 9.4
14:00 238.0 F 11.3
16:00 238.2 F 10.5
18:00 236.3 F 8.7
20:00 247.6 F 14.5
22:00 258.3 F 20.5
24:00 268.4 F 26.9
26:00 271.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 273.1 F 30.2
2:00 271.6 F 29.5
3:00 272.5 F 30.4
4:00 272.9 F 30.0

Min 271.2 F 29.5
Max 273.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 240.2 F 0.0
10:00 226.0 F 0.0
15:00 217.7 F 0.0
20:00 206.8 F 0.0
25:00 207.9 F 0.0
30:00 196.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:04:52
CYCLE COMPLETE

03/21/2025 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____