

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5308

Name: RITTER 1  
Sterilizer ID: V2178198  
02/03/2025 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.0 F 0.0  
2:00 138.7 F 2.1  
4:00 153.3 F 5.4  
6:00 176.9 F 8.7  
8:00 194.6 F 12.5  
10:00 206.8 F 3.6  
12:00 220.6 F 7.4  
14:00 233.4 F 11.5  
16:00 232.1 F 7.2  
18:00 243.5 F 12.5  
20:00 240.1 F 10.3  
22:00 252.0 F 16.1  
24:00 262.3 F 22.0  
26:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.4 F 28.8  
2:00 271.8 F 28.9  
3:00 271.5 F 28.4  
4:00 272.2 F 28.8

Min 271.3 F 28.3  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 213.6 F 0.0  
10:00 193.4 F 0.0  
15:00 242.6 F 0.0  
20:00 249.6 F 0.0  
25:00 216.2 F 0.0  
30:00 194.8 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 26:24  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:05:01  
CYCLE COMPLETE

02/03/2025 12:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_