

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001645  
15:04 05/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 2:56  
STERILIZING 3:59  
136.0 C 220kPa 3:59  
136.4 C 224kPa 4:29  
136.2 C 223kPa 4:59  
136.3 C 224kPa 5:29  
136.2 C 223kPa 5:59  
136.4 C 225kPa 6:29  
136.4 C 223kPa 6:59  
136.4 C 224kPa 7:29  
136.4 C 224kPa 7:29  
VENTING 7:29  
AIR DRYING 8:22  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
BB611396A4B881A6  
-----