
Midmark M11 - v1.0.0
Total Cycles: 1942

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2022 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	73.7 F	0.0
2:00	76.3 F	0.0
4:00	109.6 F	0.0
6:00	144.5 F	0.0
8:00	166.6 F	0.0
10:00	193.5 F	0.0
12:00	209.0 F	0.0
14:00	212.5 F	0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 213F
CHAMBER PRESS 0.1PSI
** ITEMS NOT STERILE **

05/18/2022 07:44 AM