
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004056
13:53 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:49
STERILIZING 4:07
136.4 C 223kPa 4:07
136.1 C 223kPa 4:37
136.3 C 225kPa 5:07
136.3 C 223kPa 5:37
136.2 C 221kPa 6:07
136.4 C 225kPa 6:37
136.4 C 223kPa 7:07
136.4 C 225kPa 7:37
136.4 C 225kPa 7:37
VENTING 7:37
AIR DRYING 8:31
CYCLE COMPLETE 8:36

Digital Signature #
C082381C6A0A538D
