
Midmark M11 - v1.0.5
Total Cycles: 4305

Name: RITTER2
Sterilizer ID: V2313228
01/28/2025 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.2 F 0.0
2:00 131.1 F 1.6
4:00 139.6 F 4.7
6:00 164.6 F 8.5
8:00 187.8 F 12.5
10:00 203.6 F 5.1
12:00 218.9 F 9.0
14:00 234.5 F 13.5
16:00 234.1 F 10.5
18:00 233.8 F 8.7
20:00 246.0 F 14.5
22:00 257.2 F 20.5
24:00 267.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 273.1 F 30.6
2:00 272.0 F 29.8
3:00 272.3 F 30.5
4:00 272.8 F 30.3

Min 271.6 F 29.7
Max 273.1 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 178.7 F 0.0
10:00 203.8 F 0.0
15:00 204.6 F 0.0
20:00 204.9 F 0.0
25:00 201.9 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:50
DRYING: 26:36
TOTAL CYCLE: 01:00:53

01/28/2025 03:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____