
Midmark M11 - v1.0.0
Total Cycles: 3604

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 201.1 F 4.5
4:00 204.9 F 9.0
6:00 223.1 F 7.8
8:00 231.6 F 9.1
10:00 239.6 F 9.8
12:00 242.3 F 12.2
14:00 241.1 F 10.9
16:00 254.3 F 17.7
18:00 266.0 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.8 F 28.7
3:00 272.3 F 28.9
4:00 271.6 F 28.5

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 221.6 F 0.0
10:00 232.0 F 0.0
15:00 230.1 F 0.0
20:00 227.6 F 0.0
25:00 221.9 F 0.0
30:00 226.5 F 0.0

Drying Complete

FILLING: 2:30
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

08/01/2023 04:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____