
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002516
13:02 25/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:18
STERILIZING 4:20
136.0 C 220kPa 4:20
136.4 C 222kPa 4:50
136.4 C 224kPa 5:20
136.1 C 219kPa 5:50
136.4 C 223kPa 6:20
136.2 C 222kPa 6:50
136.3 C 224kPa 7:20
136.4 C 223kPa 7:50
136.4 C 223kPa 7:50
VENTING 7:50
AIR DRYING 8:46
CYCLE COMPLETE 8:47

Digital Signature #
A2E1FE86A6537ED8
