
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002304
10:21 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:06
136.4 C 222kPa 4:06
136.3 C 225kPa 4:37
136.3 C 222kPa 5:07
136.3 C 223kPa 5:37
136.3 C 224kPa 6:07
136.4 C 223kPa 6:37
136.2 C 222kPa 7:07
136.3 C 224kPa 7:37
136.3 C 224kPa 7:37
VENTING 7:37
AIR DRYING 8:27
CYCLE COMPLETE 8:33

Digital Signature #
60DE848842D817C3
