

-----  
Midmark M11 - v1.0.5  
Total Cycles: 288

Name: RITTER2  
Sterilizer ID: V2313228  
08/19/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.9 F 0.0  
2:00 178.4 F 3.0  
4:00 202.4 F 8.5  
6:00 221.7 F 8.1  
8:00 235.8 F 11.0  
10:00 235.6 F 9.2  
12:00 238.3 F 9.6  
14:00 250.6 F 15.8  
16:00 263.2 F 23.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 271.9 F 28.7  
2:00 272.3 F 29.1  
3:00 272.1 F 29.5  
4:00 271.7 F 28.6  
  
Min 271.5 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 224.6 F 0.0  
10:00 225.0 F 0.0  
15:00 204.5 F 0.0  
Dry Cycle Aborted

FILLING: 2:57  
HEATING: 17:60  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 16:09  
TOTAL CYCLE: 00:44:10

08/19/2021 09:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02