
Midmark M11 - v1.0.0
Total Cycles: 1072

Name: RITTER 1
Sterilizer ID: V2178198
08/05/2021 06:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 79.5 F 0.0
4:00 105.1 F 0.0
6:00 122.4 F 0.1
8:00 148.7 F 0.1
10:00 181.0 F 0.2
12:00 203.2 F 2.7
14:00 214.1 F 5.9
16:00 222.7 F 8.8
18:00 231.6 F 12.3
20:00 230.8 F 7.5
22:00 241.9 F 12.4
24:00 238.6 F 9.8
26:00 237.3 F 8.6
28:00 249.0 F 14.5
30:00 260.4 F 20.8
32:00 270.6 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.6
2:00 271.5 F 28.0
3:00 272.7 F 28.7
4:00 271.5 F 27.9

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 217.4 F 0.0
10:00 193.0 F 0.0
15:00 190.9 F 0.0
20:00 187.5 F 0.0
25:00 191.1 F 0.0
30:00 207.3 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 32:49
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:51
CYCLE COMPLETE

08/05/2021 07:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____