
Midmark M11 - v1.0.5
Total Cycles: 903

Name: RITTER2
Sterilizer ID: V2313228
03/24/2022 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 122.7 F 0.0
4:00 151.4 F 0.3
6:00 184.4 F 2.8
8:00 207.6 F 6.9
10:00 220.1 F 12.0
12:00 229.7 F 8.1
14:00 239.9 F 13.5
16:00 241.2 F 11.9
18:00 240.8 F 11.2
20:00 254.4 F 18.1
22:00 266.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.3
2:00 271.6 F 29.3
3:00 272.2 F 29.2
4:00 271.6 F 29.3

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 232.5 F 0.0
10:00 186.3 F 0.0
15:00 186.5 F 0.0
20:00 185.8 F 0.0
25:00 186.7 F 0.0
30:00 187.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

03/25/2022 07:02 AM
ACCEPTED: X
REJECTED:
Operator: U02