
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005850
9:03 27/06/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:49
136.4 C 223kPa 4:49
136.3 C 223kPa 5:19
136.3 C 223kPa 5:49
136.4 C 221kPa 6:19
136.1 C 217kPa 6:49
136.3 C 222kPa 7:19
136.1 C 220kPa 7:49
136.4 C 221kPa 8:19
136.4 C 221kPa 8:19
VENTING 8:19
AIR DRYING 9:12
CYCLE COMPLETE 9:30

Digital Signature #
419482BCB79808FA
