
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001951
7:24 07/07/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:55
STERILIZING 2:57
136.5 C 222kPa 2:57
136.1 C 222kPa 3:27
136.1 C 222kPa 3:57
136.3 C 221kPa 4:27
136.3 C 219kPa 4:57
136.3 C 223kPa 5:27
136.3 C 223kPa 5:57
136.4 C 222kPa 6:27
136.4 C 222kPa 6:27
VENTING 6:27
AIR DRYING 7:19
CYCLE COMPLETE 7:26

Digital Signature #
67A16C97F69E51E3
