
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001155
16:44 06/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:55
STERILIZING 5:36
136.4 C 223kPa 5:36
136.4 C 225kPa 6:06
136.3 C 223kPa 6:36
136.2 C 224kPa 7:06
136.3 C 224kPa 7:36
136.4 C 225kPa 8:06
136.4 C 224kPa 8:36
136.4 C 224kPa 9:06
136.4 C 224kPa 9:06
VENTING 9:06
AIR DRYING 10:06
CYCLE COMPLETE 12:05

Digital Signature #
F6947B9C655361BE
