
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000104
13:58 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:22
PRESSURIZING 4:22
STERILIZING 5:34
136.0 C 222kPa 5:34
136.2 C 224kPa 6:04
136.4 C 226kPa 6:34
136.4 C 226kPa 7:04
136.4 C 226kPa 7:34
136.4 C 226kPa 8:04
136.4 C 225kPa 8:34
136.2 C 224kPa 9:04
136.2 C 224kPa 9:04
VENTING 9:04
AIR DRYING 10:06
CYCLE COMPLETE 20:49

Digital Signature #
55F40B206398D798
